



ORE RESEARCH & EXPLORATION P/L ABN 28 006 859 856  
37A Hosie Street · Bayswater North · VIC 3153 · AUSTRALIA  
☎ 61 3 9729 0333 ☎ 61 3 9729 8338  
📧 info@ore.com.au 🌐 www.ore.com.au

**CERTIFICATE OF ANALYSIS FOR**

**High Sulphidation Epithermal Ag-Cu-Au Ore**

**(Mt Carlton, Queensland, Australia)**

**CERTIFIED REFERENCE MATERIAL**

**OREAS 601b**

**Summary Statistics for Key Analytes.**


Constituent	Certified Value	1SD	95% Confidence Limits		95% Tolerance Limits	
			Low	High	Low	High
<b>Pb Fire Assay</b>						
Au, Gold (ppm)	0.775	0.021	0.767	0.782	0.770*	0.779*
<b>4-Acid Digestion</b>						
Ag, Silver (ppm)	50.1	1.74	49.4	50.9	49.0	51.3
Cu, Copper (wt.%)	0.101	0.002	0.100	0.102	0.099	0.102

SI unit equivalents: ppm, parts per million  $\equiv$  mg/kg  $\equiv$   $\mu$ g/g  $\equiv$  0.0001 wt.%  $\equiv$  1000 ppb, parts per billion.

\*Gold Tolerance Limits for typical 25-50g fire assay method is determined from 20 x 85mg INAA results and the Sampling Constant (Ingamells & Switzer, 1973).

Note 1: intervals may appear asymmetric due to rounding.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.



Document: COA-1399-OREAS601b-R2

(Template:BUP-70-10-01 Rev:2.0)

20-June-2019

## TABLE OF CONTENTS

INTRODUCTION .....	7
SOURCE MATERIAL.....	7
COMMINUTION AND HOMOGENISATION PROCEDURES .....	8
PHYSICAL PROPERTIES .....	8
ANALYTICAL PROGRAM.....	8
STATISTICAL ANALYSIS.....	9
Homogeneity Evaluation .....	11
Performance Gates .....	13
PARTICIPATING LABORATORIES.....	16
PREPARER AND SUPPLIER.....	20
METROLOGICAL TRACEABILITY .....	20
COMMUTABILITY .....	20
INTENDED USE .....	21
STABILITY AND STORAGE INSTRUCTIONS .....	21
INSTRUCTIONS FOR CORRECT USE.....	21
HANDLING INSTRUCTIONS.....	21
LEGAL NOTICE.....	21
DOCUMENT HISTORY .....	21
QMS ACCREDITED .....	22
CERTIFYING OFFICER.....	22
REFERENCES .....	22

## LIST OF TABLES

Table 1. Certified Values, SDs, 95% Confidence & Tolerance Limits for OREAS 601b.....	3
Table 2. Indicative Values for OREAS 601b. ....	6
Table 3. Physical properties of OREAS 601b. ....	8
Table 4. Neutron Activation Analysis of Au (in ppm) on 20 x 85mg subsamples .....	11
Table 5. Performance Gates for OREAS 601b. ....	13

## LIST OF FIGURES

Figure 1. 'Au' (ppm) by Pb Fire Assay in OREAS 601b .....	17
Figure 2. 'Ag' (ppm) by 4-acid digestion in OREAS 601b .....	18
Figure 3. 'Cu' (wt.%) by 4-acid digestion in OREAS 601b .....	19

**Table 1. Certified Values, SDs, 95% Confidence & Tolerance Limits for OREAS 601b.**

Constituent	Certified Value	SD	95% Confidence Limits		95% Tolerance Limits	
			Low	High	Low	High
<b>Pb Fire Assay</b>						
Au, Gold (ppm)	0.775	0.021	0.767	0.782	0.770*	0.779*
<b>Aqua Regia Digestion (sample weights 10-50g)</b>						
Au, Gold (ppm)	0.761	0.033	0.741	0.780	0.755*	0.766*
<b>Infrared Combustion</b>						
S, Sulphur (wt.%)	1.49	0.072	1.46	1.52	1.46	1.52
<b>4-Acid Digestion</b>						
Ag, Silver (ppm)	50.1	1.74	49.4	50.9	49.0	51.3
Al, Aluminium (wt.%)	6.63	0.223	6.53	6.73	6.53	6.73
As, Arsenic (ppm)	284	18	275	293	278	291
Be, Beryllium (ppm)	2.24	0.33	2.08	2.41	2.12	2.37
Bi, Bismuth (ppm)	18.0	1.14	17.4	18.6	17.4	18.6
Ca, Calcium (wt.%)	0.887	0.044	0.867	0.906	0.871	0.902
Cd, Cadmium (ppm)	2.05	0.111	2.00	2.10	1.96	2.13
Ce, Cerium (ppm)	70	8	65	75	68	72
Co, Cobalt (ppm)	2.97	0.184	2.89	3.06	2.82	3.13
Cr, Chromium (ppm)	23.7	3.0	22.3	25.1	22.5	24.9
Cs, Cesium (ppm)	4.88	0.185	4.79	4.97	4.73	5.03
Cu, Copper (wt.%)	0.101	0.002	0.100	0.102	0.099	0.102
Dy, Dysprosium (ppm)	2.54	0.160	2.38	2.71	2.38	2.71
Er, Erbium (ppm)	0.80	0.061	0.74	0.86	IND	IND
Eu, Europium (ppm)	0.97	0.11	0.83	1.11	IND	IND
Fe, Iron (wt.%)	2.29	0.081	2.25	2.32	2.24	2.33
Ga, Gallium (ppm)	23.4	1.78	22.5	24.2	22.7	24.1
Gd, Gadolinium (ppm)	4.14	0.311	3.78	4.50	3.91	4.36
Ge, Germanium (ppm)	0.15	0.04	0.10	0.20	IND	IND
Hf, Hafnium (ppm)	5.09	0.296	4.94	5.25	4.95	5.23
Ho, Holmium (ppm)	0.38	0.04	0.35	0.42	IND	IND
In, Indium (ppm)	0.47	0.026	0.46	0.48	0.43	0.51
K, Potassium (wt.%)	2.41	0.075	2.38	2.44	2.36	2.46
La, Lanthanum (ppm)	33.5	2.52	32.2	34.8	32.5	34.5
Li, Lithium (ppm)	22.6	1.98	21.7	23.4	21.9	23.3
Lu, Lutetium (ppb)	73.1	13.3	58.0	88.2	IND	IND
Mg, Magnesium (ppm)	996	66	968	1024	976	1016
Mn, Manganese (ppm)	222	8	219	225	216	227
Mo, Molybdenum (ppm)	5.22	0.471	5.00	5.44	4.96	5.49

SI unit equivalents: ppm, parts per million  $\equiv$  mg/kg  $\equiv$   $\mu$ g/g  $\equiv$  0.0001 wt.%  $\equiv$  1000 ppb, parts per billion.

\*Gold Tolerance Limits for typical 25-50g fire assay and 15-40g aqua regia digestion methods are determined from 20 x 85mg INAA results and the Sampling Constant (Ingamells & Switzer, 1973).

Note 1: intervals may appear asymmetric due to rounding.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.

Table 1 continued.

Constituent	Certified Value	SD	95% Confidence Limits		95% Tolerance Limits	
			Low	High	Low	High
<b>4-Acid Digestion continued</b>						
Na, Sodium (wt.%)	1.90	0.090	1.86	1.94	1.86	1.94
Nb, Niobium (ppm)	14.4	1.04	13.9	14.9	13.9	14.9
Nd, Neodymium (ppm)	28.5	2.19	25.6	31.4	27.7	29.4
Ni, Nickel (ppm)	6.54	0.67	6.16	6.91	6.06	7.01
P, Phosphorus (ppm)	292	15	286	297	282	301
Pb, Lead (ppm)	318	16	311	325	310	326
Pr, Praseodymium (ppm)	8.55	0.572	7.81	9.29	8.27	8.84
Rb, Rubidium (ppm)	98	3.5	96	99	95	100
S, Sulphur (wt.%)	1.50	0.036	1.48	1.51	1.47	1.52
Sb, Antimony (ppm)	22.9	2.09	21.9	23.9	22.2	23.6
Sc, Scandium (ppm)	3.77	0.239	3.64	3.90	3.60	3.94
Se, Selenium (ppm)	10.6	1.1	10.0	11.1	10.1	11.0
Sm, Samarium (ppm)	5.77	0.432	5.24	6.30	5.56	5.98
Sn, Tin (ppm)	3.36	0.177	3.26	3.45	3.17	3.54
Sr, Strontium (ppm)	241	12	235	247	237	245
Ta, Tantalum (ppm)	1.11	0.076	1.07	1.15	1.08	1.14
Tb, Terbium (ppm)	0.52	0.07	0.44	0.60	0.50	0.54
Te, Tellurium (ppm)	12.6	0.96	12.1	13.1	12.0	13.2
Th, Thorium (ppm)	11.9	0.82	11.5	12.3	11.6	12.2
Ti, Titanium (wt.%)	0.135	0.005	0.132	0.137	0.132	0.137
Tl, Thallium (ppm)	1.44	0.092	1.39	1.49	1.39	1.49
Tm, Thulium (ppb)	< 100	IND	IND	IND	IND	IND
U, Uranium (ppm)	4.64	0.216	4.53	4.75	4.48	4.80
V, Vanadium (ppm)	12.1	0.88	11.7	12.6	11.8	12.4
W, Tungsten (ppm)	6.13	0.339	6.00	6.27	5.86	6.41
Y, Yttrium (ppm)	11.1	0.51	10.9	11.3	10.8	11.5
Yb, Ytterbium (ppm)	0.54	0.06	0.48	0.61	0.47	0.62
Zn, Zinc (ppm)	318	6	316	320	312	325
Zr, Zirconium (ppm)	186	11	181	191	182	190
<b>Aqua Regia Digestion</b>						
Ag, Silver (ppm)	50.0	2.86	48.8	51.2	49.0	50.9
Al, Aluminium (wt.%)	0.630	0.035	0.613	0.646	0.611	0.648
As, Arsenic (ppm)	276	17	269	284	270	282
Be, Beryllium (ppm)	0.47	0.041	0.46	0.49	0.44	0.51
Bi, Bismuth (ppm)	18.0	1.58	17.3	18.8	17.4	18.7
Ca, Calcium (wt.%)	0.578	0.030	0.565	0.591	0.565	0.591

SI unit equivalents: ppm, parts per million  $\equiv$  mg/kg  $\equiv$   $\mu$ g/g  $\equiv$  0.0001 wt.%  $\equiv$  1000 ppb, parts per billion.

Note 1: intervals may appear asymmetric due to rounding.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.

Table 1 continued.

Constituent	Certified Value	SD	95% Confidence Limits		95% Tolerance Limits	
			Low	High	Low	High
<b>Aqua Regia Digestion continued</b>						
Cd, Cadmium (ppm)	2.04	0.177	1.95	2.13	1.94	2.14
Ce, Cerium (ppm)	38.5	2.32	37.2	39.9	37.3	39.8
Co, Cobalt (ppm)	2.55	0.121	2.50	2.60	2.42	2.69
Cr, Chromium (ppm)	24.6	2.26	23.7	25.5	23.3	25.9
Cs, Cesium (ppm)	1.15	0.092	1.09	1.20	1.10	1.19
Cu, Copper (wt.%)	0.101	0.003	0.100	0.102	0.099	0.103
Fe, Iron (wt.%)	1.94	0.114	1.89	1.99	1.90	1.98
Ga, Gallium (ppm)	3.77	0.189	3.67	3.87	3.62	3.92
Ge, Germanium (ppm)	0.084	0.024	0.062	0.106	IND	IND
Hf, Hafnium (ppm)	1.11	0.107	1.04	1.17	1.06	1.16
Hg, Mercury (ppm)	0.20	0.02	0.18	0.21	0.18	0.22
In, Indium (ppm)	0.42	0.016	0.42	0.43	0.40	0.44
K, Potassium (wt.%)	0.246	0.019	0.236	0.256	0.237	0.255
La, Lanthanum (ppm)	19.9	1.69	19.2	20.7	19.1	20.8
Li, Lithium (ppm)	7.78	0.695	7.42	8.13	7.49	8.06
Mg, Magnesium (ppm)	411	20	400	423	392	430
Mn, Manganese (ppm)	192	8	188	195	189	195
Mo, Molybdenum (ppm)	4.83	0.281	4.73	4.94	4.60	5.07
Na, Sodium (wt.%)	0.067	0.008	0.063	0.070	0.065	0.069
Ni, Nickel (ppm)	6.39	0.419	6.21	6.58	6.09	6.70
P, Phosphorus (ppm)	178	8	174	181	170	186
Pb, Lead (ppm)	234	14	227	240	229	238
Rb, Rubidium (ppm)	11.9	0.71	11.5	12.3	11.6	12.3
S, Sulphur (wt.%)	0.807	0.038	0.789	0.826	0.790	0.825
Sb, Antimony (ppm)	18.0	2.2	17.0	19.1	17.6	18.5
Sc, Scandium (ppm)	1.00	0.13	0.94	1.07	0.90	1.10
Se, Selenium (ppm)	10.0	1.2	9.4	10.6	9.4	10.6
Sn, Tin (ppm)	1.19	0.112	1.13	1.26	1.13	1.26
Sr, Strontium (ppm)	33.2	2.95	31.8	34.7	32.0	34.5
Tb, Terbium (ppm)	0.28	0.06	0.21	0.35	0.26	0.29
Te, Tellurium (ppm)	13.0	0.83	12.6	13.4	12.6	13.4
Th, Thorium (ppm)	6.96	0.73	6.56	7.36	6.72	7.20
Ti, Titanium (ppm)	220	22	207	233	210	230
Tl, Thallium (ppm)	1.08	0.084	1.03	1.12	1.04	1.12
U, Uranium (ppm)	2.22	0.119	2.15	2.30	2.14	2.31
V, Vanadium (ppm)	3.83	0.51	3.59	4.06	3.65	4.00

SI unit equivalents: ppm, parts per million  $\equiv$  mg/kg  $\equiv$   $\mu$ g/g  $\equiv$  0.0001 wt.%  $\equiv$  1000 ppb, parts per billion.

Note 1: intervals may appear asymmetric due to rounding.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.

Table 1 continued.

Constituent	Certified Value	SD	95% Confidence Limits		95% Tolerance Limits	
			Low	High	Low	High
<b>Aqua Regia Digestion continued</b>						
W, Tungsten (ppm)	1.86	0.184	1.76	1.96	1.70	2.01
Y, Yttrium (ppm)	5.34	0.528	5.08	5.60	5.15	5.52
Yb, Ytterbium (ppm)	0.20	0.020	0.18	0.23	0.18	0.22
Zn, Zinc (ppm)	267	7	263	270	263	271
Zr, Zirconium (ppm)	38.3	5.5	35.6	40.9	36.9	39.6

SI unit equivalents: ppm, parts per million  $\equiv$  mg/kg  $\equiv$   $\mu$ g/g  $\equiv$  0.0001 wt.%  $\equiv$  1000 ppb, parts per billion.

Note 1: intervals may appear asymmetric due to rounding.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.

Table 2. Indicative Values for OREAS 601b.

Constituent	Unit	Value	Constituent	Unit	Value	Constituent	Unit	Value
<b>Pb Fire Assay</b>								
Ag	ppm	53.7	Pd	ppb	< 3	Pt	ppb	< 5
<b>Infrared Combustion</b>								
C	wt.%	0.157						
<b>4-Acid Digestion</b>								
Au	ppm	0.752	Ir	ppb	< 5	Re	ppb	< 2
Ba	ppm	805	Pd	ppb	< 50	Rh	ppb	< 10
Hg	ppm	< 2	Pt	ppb	< 10	Ru	ppb	< 10
<b>Aqua Regia Digestion</b>								
B	ppm	< 10	Ir	ppb	< 5	Re	ppb	< 1
Ba	ppm	280	Lu	ppb	28.6	Rh	ppb	< 10
Dy	ppm	1.29	Nb	ppm	1.13	Ru	ppb	< 10
Er	ppm	0.36	Nd	ppm	16.3	Sm	ppm	3.04
Eu	ppm	0.53	Pd	ppb	< 10	Ta	ppm	< 0.01
Gd	ppm	2.33	Pr	ppm	4.56	Tm	ppb	40.0
Ho	ppm	0.18	Pt	ppb	< 5			
<b>Borate Fusion XRF</b>								
Al <sub>2</sub> O <sub>3</sub>	wt.%	12.91	MgO	wt.%	0.175	SiO <sub>2</sub>	wt.%	71.32
CaO	wt.%	1.22	MnO	wt.%	0.030	SO <sub>3</sub>	wt.%	3.76
Fe <sub>2</sub> O <sub>3</sub>	wt.%	3.34	Na <sub>2</sub> O	wt.%	2.60	TiO <sub>2</sub>	wt.%	0.210
K <sub>2</sub> O	wt.%	2.93	P <sub>2</sub> O <sub>5</sub>	wt.%	0.069			
<b>Thermogravimetry</b>								
H <sub>2</sub> O-	wt.%	0.397	LOI <sup>1000</sup>	wt.%	4.42			
<b>Laser Ablation ICP-MS</b>								
Ag	ppm	51.1	Hf	ppm	5.94	Sm	ppm	5.49
As	ppm	271	Ho	ppm	0.39	Sn	ppm	3.90
Ba	ppm	4145	In	ppm	0.45	Sr	ppm	244
Be	ppm	2.70	La	ppm	37.3	Ta	ppm	1.14
Bi	ppm	19.3	Lu	ppb	80.0	Tb	ppm	0.58
Cd	ppm	2.25	Mn	ppm	218	Te	ppm	13.3

SI unit equivalents: ppm, parts per million  $\equiv$  mg/kg  $\equiv$   $\mu$ g/g  $\equiv$  0.0001 wt.%  $\equiv$  1000 ppb, parts per billion.

Note: the number of significant figures reported is not a reflection of the level of certainty of stated values. They are instead an artefact of ORE's in-house CRM-specific LIMS.

**Table 2 continued.**

Constituent	Unit	Value	Constituent	Unit	Value	Constituent	Unit	Value
<b>Laser Ablation ICP-MS continued</b>								
Ce	ppm	71	Mo	ppm	5.30	Th	ppm	12.4
Co	ppm	3.00	Nb	ppm	15.0	Ti	wt.%	0.137
Cr	ppm	28.5	Nd	ppm	29.6	Tl	ppm	1.70
Cs	ppm	4.88	Ni	ppm	9.00	Tm	ppb	100
Cu	wt.%	0.095	Pb	ppm	321	U	ppm	4.70
Dy	ppm	2.69	Pr	ppm	7.92	V	ppm	11.8
Er	ppm	0.87	Rb	ppm	97	W	ppm	6.00
Eu	ppm	0.96	Re	ppb	25.0	Y	ppm	11.8
Ga	ppm	22.5	Sb	ppm	25.6	Yb	ppm	0.56
Gd	ppm	4.00	Sc	ppm	5.85	Zn	ppm	288
Ge	ppm	1.78	Se	ppm	< 5	Zr	ppm	222

SI unit equivalents: ppm, parts per million  $\equiv$  mg/kg  $\equiv$   $\mu$ g/g  $\equiv$  0.0001 wt.%  $\equiv$  1000 ppb, parts per billion.

Note: the number of significant figures reported is not a reflection of the level of certainty of stated values. They are instead an artefact of ORE's in-house CRM-specific LIMS.

## INTRODUCTION

OREAS reference materials are intended to provide a low cost method of evaluating and improving the quality of analysis of geological samples. To the geologist they provide a means of implementing quality control in analytical data sets generated in exploration from the grass roots level through to prospect evaluation, and in grade control at mining operations. To the analyst they provide an effective means of calibrating analytical equipment, assessing new techniques and routinely monitoring in-house procedures.

OREAS reference materials enable users to successfully achieve process control of these tasks because the observed variance from repeated analysis has its origin almost exclusively in the analytical process rather than the reference material itself.

## SOURCE MATERIAL

OREAS 601b was prepared from a blend of silver-copper-gold bearing ores from Evolution Mining's Mount Carlton Operation in Queensland, Australia and argillic rhyodacite waste rock sourced from a quarry east of Melbourne, Australia. A small amount of chalcocite-rich ore from the Sepon Mine in Laos was also added to help achieve the desired copper grade.

The mineralisation assemblage at Mount Carlton consists of pyrite, enargite/tennantite, tetrahedrite, digenite, covellite, sphalerite, galena, alunite, dickite, kaolinite and vuggy silica, hosted in advanced argillic altered rhyodacite containing sulphur-salts.

OREAS 601b is one of a suite of six CRMs developed from Mount Carlton ores ranging in grades from 25 -1015ppm Ag, 0.2 -1.7ppm Au and 0.05 - 5.0% Cu.

## COMMINUTION AND HOMOGENISATION PROCEDURES

The material constituting OREAS 601b was prepared in the following manner:

- Drying of ore materials (sulphide-rich) to constant mass at 85°C;
- Drying of rhyodacite waste rock to constant mass at 105°C;
- Crushing and milling of the ore materials to 100% minus 30 microns;
- Crushing and milling of the rhyodacite waste rock to 98% minus 75 microns;
- Blending in appropriate proportions to achieve the desired grades;
- Packaging under nitrogen in 10g and 60g units in laminated foil pouches and 500g units in plastic jars.

## PHYSICAL PROPERTIES

OREAS 601b was tested at ORE Research & Exploration Pty Ltd's onsite laboratory for various physical properties. Table 3 presents these findings which should be used for informational purposes only.

**Table 3. Physical properties of OREAS 601b.**

CRM Name	Bulk Density (g/L)	Moisture%	Munsell Notation <sup>‡</sup>	Munsell Color <sup>‡</sup>
OREAS 601b	718	0.44	N8	Very Light Gray

<sup>‡</sup>The Munsell Rock Color Chart helps geologists and archeologists communicate with color more effectively by cross-referencing ISCC-NBS color names with unique Munsell alpha-numeric color notations for rock color samples.

## ANALYTICAL PROGRAM

Twenty five commercial analytical laboratories participated in the program to certify the elements reported in Table 1. The following methods were employed:

- Gold via 25-50g fire assay with AAS finish (14 laboratories) and ICP-OES (10 laboratories) finish;
- Gold via 10-40g aqua regia digestion with ICP-MS finish (10 laboratories) and AAS (2 laboratories) finish;
- Sulphur by infra-red combustion analysis (19 laboratories);
- 4-Acid digestion for full elemental suite ICP-OES/MS finish (up to 22 laboratories depending on the element).
- Aqua regia digestion for full elemental suite ICP-OES finish (up to 22 laboratories depending on the element) and AAS finish (1 laboratory);
- Gold by instrumental neutron activation analysis (INAA) on 20 x 85mg subsamples to confirm homogeneity (ANSTO, Lucas Heights).

It is important to note that in the analytical industry there is no standardisation of the aqua regia digestion process. Aqua regia is a partial empirical digest and differences in recoveries for various analytes are commonplace. These are caused by variations in the digest conditions which can include the ratio of nitric to hydrochloric acids, acid strength, temperatures, leach times and secondary digestions. Recoveries for sulphide-hosted base



metal sulphides approach total values, however, other analytes, in particular the lithophile elements, show greater sensitivity to method parameters. This can result in lack of consensus in an inter-laboratory certification program for these elements.

The approach applied here is to report certified values in those instances where reasonable agreement exists amongst a majority of participating laboratories. The results of specific laboratories may differ significantly from the certified values, but will, nonetheless, be valid and reproducible in the context of the specifics of the aqua regia method in use. Users of this reference material should, therefore, be mindful of this limitation when applying the certified values in a quality control program.

For the round robin program twenty 1kg test units were taken at predetermined intervals during the bagging stage, immediately following homogenisation and are considered representative of the entire prepared batch. Six 100g pulp samples were submitted to each laboratory for analysis received by each laboratory were obtained by taking two 100g samples from each of three separate 1kg test units. This format enabled nested ANOVA treatment of the results to evaluate homogeneity, i.e. to ascertain whether between-unit variance is greater than within-unit variance.

Table 1 presents the 108 certified values together with their associated 1SD's, 95% confidence and tolerance limits, Table 2 shows 97 indicative values for major and trace element composition. Gold homogeneity has been evaluated and confirmed by instrumental neutron activation analysis (INAA) on twenty ~85mg sample portions (see Table 4 below) and by a nested ANOVA program for both fire assay and aqua regia digestion (see '**nested ANOVA**' section).

Table 5 provides performance gate intervals for the certified values based on their pooled 1SD's. Tabulated results of all elements together with uncorrected means, medians, standard deviations, relative standard deviations and per cent deviation of lab means from the corrected mean of means (PDM<sup>3</sup>) are presented in the detailed certification data for this CRM (**OREAS 601b DataPack-1.0.190219\_134415.xlsx**).

Results are also presented in scatter plots for gold by fire assay, silver by 4-acid digestion and copper by 4-acid digestion (Figures 1 to 3, respectively) together with  $\pm 3SD$  (magenta) and  $\pm 5\%$  (yellow) control lines and certified value (green line). Accepted individual results are coloured blue and individual and dataset outliers are identified in red and violet, respectively.

## STATISTICAL ANALYSIS

**Certified Values, Confidence Limits, Standard Deviations and Tolerance Limits** (Table 1) have been determined for each analyte following removal of individual, laboratory dataset (batch) and 3SD outliers (single iteration).

For individual outliers within a laboratory batch the z-score test is used in combination with a second method that determines the per cent deviation of the individual value from the batch median. Outliers in general are selected on the basis of z-scores  $> 2.5$  and with per cent deviations (i)  $> 3$  and (ii) more than three times the average absolute per cent deviation for the batch. In certain instances statistician's prerogative has been employed in discriminating outliers.

Each laboratory data set mean is tested for outlying status based on z-score discrimination and rejected if  $> 2.5$ . After individual and laboratory data set (batch) outliers have been

eliminated a non-iterative 3 standard deviation filter is applied, with those values lying outside this window also relegated to outlying status.

**Certified Values** are the means of accepted laboratory means after outlier filtering. The INAA data (see Table 4) is omitted from determination of the certified value for Au and is used solely for the calculation of Tolerance Limits and homogeneity evaluation of OREAS 601b.

**95% Confidence Limits** are inversely proportional to the number of participating laboratories and inter-laboratory agreement. It is a measure of the reliability of the certified value. A 95% confidence interval indicates a 95% probability that the true value of the analyte under consideration lies between the upper and lower limits. *95% Confidence Limits should not be used as control limits for laboratory performance.*

**Indicative (uncertified) values** (Table 2) are provided for the major and trace elements determined by borate fusion XRF ( $\text{Al}_2\text{O}_3$  to  $\text{TiO}_2$ ), laser ablation with ICP-MS (Ag to Zr), LOI at  $1000^\circ\text{C}$  and C by infrared combustion furnace and are the means of duplicate assays from Bureau Veritas, Perth. Additional indicative values by other analytical methods are present where the number of laboratories reporting a particular analyte is insufficient ( $< 5$ ) to support certification or where inter-laboratory consensus is poor.

**Standard Deviation** values (1SDs) are reported in Table 1. They provide an indication of a level of performance that might reasonably be expected from a laboratory being monitored by this CRM in a QA/QC program. They take into account errors attributable to measurement uncertainty and CRM variability. For an effective CRM the contribution of the latter should be negligible in comparison to measurement errors. The Standard Deviation values include all sources of measurement uncertainty: between-lab variance, within-run variance (precision errors) and CRM variability.

The SD for each analyte's certified value is calculated from the same filtered data set used to determine the certified value, i.e. after removal of all individual, lab dataset (batch) and 3SD outliers (single iteration). These outliers can only be removed after the absolute homogeneity of the CRM has been independently established, i.e. the outliers must be confidently deemed to be analytical rather than arising from inhomogeneity of the CRM. **The standard deviation is then calculated for each analyte from the pooled accepted analyses generated from the certification program.**

In the application of SD's in monitoring performance it is important to note that not all laboratories function at the same level of proficiency and that different methods in use at a particular laboratory have differing levels of precision. Each laboratory has its own inherent SD (for a specific concentration level and analyte-method pair) based on the analytical process and this SD is not directly related to the round robin program.

The majority of data generated in the round robin program was produced by a selection of world class laboratories. The SD's thus generated are more constrained than those that would be produced across a randomly selected group of laboratories. To produce more generally achievable SD's the 'pooled' SD's provided in this report include inter-lab bias. This 'one size fits all' approach may require revision at the discretion of the QC manager concerned following careful scrutiny of QC control charts.

## Homogeneity Evaluation

The tolerance limits (ISO 16269:2014) shown in Table 1 were determined using an analysis of precision errors method and are considered a conservative estimate of true homogeneity. The meaning of tolerance limits may be illustrated for copper by 4-acid digestion, where 99% of the time ( $1-\alpha=0.99$ ) at least 95% of subsamples ( $\rho=0.95$ ) will have concentrations lying between 0.099 and 0.102 wt.%. Put more precisely, this means that if the same number of subsamples were taken and analysed in the same manner repeatedly, 99% of the tolerance intervals so constructed would cover at least 95% of the total population, and 1% of the tolerance intervals would cover less than 95% of the total population (ISO Guide 35). **Please note that tolerance limits pertain to the homogeneity of the CRM only and should not be used as control limits for laboratory performance.**

Table 4 below shows the INAA data determined on 20 x 85mg subsamples of OREAS 601b. An equivalent scaled version of the results is also provided to demonstrate an appreciation of what this data means if 30g fire assay determinations were undertaken without the normal measurement error associated with this methodology.

**Table 4. Neutron Activation Analysis of Au (in ppm) on 20 x 85mg subsamples showing the equivalent results scaled to a 30g sample mass typical of fire assay determination.**

Replicate	Au	Au
No	85mg actual	30g equivalent*
1	0.803	0.797
2	0.774	0.796
3	0.786	0.796
4	0.891	0.802
5	0.795	0.797
6	0.799	0.797
7	0.774	0.796
8	0.787	0.796
9	0.774	0.796
10	0.750	0.794
11	0.817	0.798
12	0.810	0.797
13	0.834	0.799
14	0.778	0.796
15	0.807	0.797
16	0.776	0.796
17	0.793	0.797
18	0.775	0.796
19	0.799	0.797
20	0.812	0.798
Mean	0.797	0.797
Median	0.794	0.797
Std Dev.	0.029	0.002
<b>Rel.Std.Dev.</b>	<b>3.69%</b>	<b>0.197%</b>

\*Results calculated for a 30g equivalent sample mass using the formula:  $x^{30g Eq} = \frac{(x^{INAA} - \bar{X}) \times RSD@30g}{RSD@85mg} + \bar{X}$

where  $x^{30g Eq}$  = equivalent result calculated for a 30g sample mass

$(x^{INAA})$  = raw INAA result at 85mg

$\bar{X}$  = mean of 85mg INAA results

The homogeneity of gold has been determined by INAA using the reduced analytical subsample method which utilises the known relationship between standard deviation and analytical subsample weight (Ingamells and Switzer, 1973). In this approach the sample aliquot is substantially reduced to a point where most of the variability in replicate assays should be due to inhomogeneity of the reference material and measurement error becomes negligible. In this instance a subsample weight of 85 milligrams was employed and the 1RSD of 0.197% was calculated for a 30g fire assay or aqua regia sample (3.96% at 85mg weights) confirms the high level of gold homogeneity in OREAS 601b.

The homogeneity of OREAS 601b has also been evaluated in a **nested ANOVA** of the round robin program. Each of the twenty-five round robin laboratories received six samples per CRM and these samples were made up of paired samples from three different, non-adjacent sampling intervals. The purpose of the ANOVA evaluation is to test that no statistically significant difference exists in the variance between-units to that of the variance within-units. This allows an assessment of homogeneity across the entire prepared batch of OREAS 601b. The test was performed using the following parameters:

- Gold fire assay – 144 samples (24 laboratories each providing analyses on 3 pairs of samples);
- Gold aqua regia digestion – 72 samples (12 laboratories each providing analyses on 3 pairs of samples);
- Null Hypothesis,  $H_0$ : Between-unit variance is no greater than within-unit variance (reject  $H_0$  if  $p$ -value < 0.05);
- Alternative Hypothesis,  $H_1$ : Between-unit variance is greater than within-unit variance.

$P$ -values are a measure of probability where values less than 0.05 indicate a greater than 95% probability that the observed differences in within-unit and between-unit variances are real. The datasets were filtered for both individual and laboratory data set (batch) outliers prior to the calculation of  $p$ -values. This process derived  $p$ -values of 0.853 for Au by fire assay and 0.062 for Au by aqua regia digestion. Both  $p$ -values are insignificant and the Null Hypothesis is retained. Additionally, none of the other certified values showed significant  $p$ -values except for Zn by aqua regia digestion ( $p$ -value = 0.012). This analyte has a low associated 1RSD of 2.79% meaning the data is well constrained. This isolated case is most likely due to random statistical probability (as there is no other supporting evidence to suspect greater between-unit variance compared with within-unit variance. The null hypothesis is therefore retained.

Please note that only results for constituents present in concentrations well above the detection levels (i.e. >20 x Lower Limit of Detection) for the various methods undertaken were considered for the objective of evaluating homogeneity. It is important to note that ANOVA is not an absolute measure of homogeneity. Rather, it establishes whether or not the analytes are distributed in a similar manner throughout the packaging run of OREAS 601b and whether the variance between two subsamples from the same unit is statistically distinguishable to the variance from two subsamples taken from any two separate units. A reference material therefore, can possess poor absolute homogeneity yet still pass a relative homogeneity test if the within-unit heterogeneity is large and similar across all units.

Based on the statistical analysis of the results of the inter-laboratory certification program it can be concluded that OREAS 601b is fit-for-purpose as a certified reference material (see 'Intended Use' below).

## Performance Gates

Table 5 shows calculated for two and three standard deviations. As a guide these intervals may be regarded as warning or rejection for multiple 2SD outliers, or rejection for individual 3SD outliers in QC monitoring, although their precise application should be at the discretion of the QC manager concerned. A second method utilises a 5% window calculated directly from the certified value.

Standard deviation is also shown in relative percent for one, two and three relative standard deviations (1RSD, 2RSD and 3RSD) to facilitate an appreciation of the magnitude of these numbers and a comparison with the 5% window. Caution should be exercised when concentration levels approach lower limits of detection of the analytical methods employed as performance gates calculated from standard deviations tend to be excessively wide whereas those determined by the 5% method are too narrow. One approach used at commercial laboratories is to set the acceptance criteria at twice the detection level (DL)  $\pm 10\%$ .

*i.e. Certified Value  $\pm 10\% \pm 2DL$  (adapted from Govett, 1983)*

**Table 5. Performance Gates for OREAS 601b.**

Constituent	Certified Value	Absolute Standard Deviations					Relative Standard Deviations			5% window	
		1SD	2SD Low	2SD High	3SD Low	3SD High	1RSD	2RSD	3RSD	Low	High
<b>Pb Fire Assay</b>											
Au, ppm	0.775	0.021	0.733	0.816	0.712	0.837	2.70%	5.41%	8.11%	0.736	0.813
<b>Aqua Regia Digestion (sample weights 10-50g)</b>											
Au, ppm	0.761	0.033	0.694	0.827	0.661	0.860	4.37%	8.74%	13.12%	0.723	0.799
<b>Infrared Combustion</b>											
S, wt. %	1.49	0.072	1.35	1.64	1.28	1.71	4.81%	9.63%	14.44%	1.42	1.57
<b>4-Acid Digestion</b>											
Ag, ppm	50.1	1.74	46.6	53.6	44.9	55.3	3.47%	6.94%	10.40%	47.6	52.6
Al, wt. %	6.63	0.223	6.19	7.08	5.96	7.30	3.37%	6.73%	10.10%	6.30	6.97
As, ppm	284	18	248	321	230	339	6.41%	12.82%	19.22%	270	298
Be, ppm	2.24	0.33	1.58	2.91	1.25	3.24	14.74%	29.49%	44.23%	2.13	2.36
Bi, ppm	18.0	1.14	15.7	20.3	14.6	21.4	6.35%	12.70%	19.04%	17.1	18.9
Ca, wt. %	0.887	0.044	0.799	0.975	0.755	1.019	4.97%	9.94%	14.91%	0.842	0.931
Cd, ppm	2.05	0.111	1.83	2.27	1.71	2.38	5.44%	10.88%	16.32%	1.95	2.15
Ce, ppm	70	8	53	87	45	95	11.83%	23.66%	35.49%	66	73
Co, ppm	2.97	0.184	2.61	3.34	2.42	3.53	6.20%	12.39%	18.59%	2.83	3.12
Cr, ppm	23.7	3.0	17.8	29.6	14.8	32.6	12.52%	25.05%	37.57%	22.5	24.9
Cs, ppm	4.88	0.185	4.51	5.25	4.33	5.44	3.80%	7.59%	11.39%	4.64	5.13
Cu, wt. %	0.101	0.002	0.096	0.105	0.094	0.108	2.26%	4.53%	6.79%	0.096	0.106
Dy, ppm	2.54	0.160	2.22	2.86	2.06	3.02	6.30%	12.61%	18.91%	2.41	2.67
Er, ppm	0.80	0.061	0.68	0.92	0.62	0.98	7.62%	15.24%	22.87%	0.76	0.84
Eu, ppm	0.97	0.11	0.75	1.19	0.64	1.30	11.21%	22.41%	33.62%	0.92	1.02
Fe, wt. %	2.29	0.081	2.12	2.45	2.04	2.53	3.55%	7.11%	10.66%	2.17	2.40
Ga, ppm	23.4	1.78	19.8	26.9	18.0	28.7	7.62%	15.24%	22.86%	22.2	24.6

SI unit equivalents: ppm, parts per million  $\equiv$  mg/kg  $\equiv$   $\mu$ g/g  $\equiv$  0.0001 wt. %  $\equiv$  1000 ppb, parts per billion.

Note 1: intervals may appear asymmetric due to rounding.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.



Table 5 continued.

Constituent	Certified Value	Absolute Standard Deviations					Relative Standard Deviations			5% window	
		1SD	2SD Low	2SD High	3SD Low	3SD High	1RSD	2RSD	3RSD	Low	High
<b>4-Acid Digestion continued</b>											
Gd, ppm	4.14	0.311	3.52	4.76	3.21	5.07	7.51%	15.01%	22.52%	3.93	4.34
Ge, ppm	0.15	0.04	0.06	0.24	0.02	0.28	29.04%	58.08%	87.12%	0.14	0.16
Hf, ppm	5.09	0.296	4.50	5.69	4.21	5.98	5.80%	11.60%	17.40%	4.84	5.35
Ho, ppm	0.38	0.04	0.30	0.47	0.26	0.51	10.89%	21.78%	32.66%	0.37	0.40
In, ppm	0.47	0.026	0.42	0.52	0.39	0.55	5.65%	11.30%	16.95%	0.45	0.49
K, wt.%	2.41	0.075	2.26	2.56	2.18	2.64	3.13%	6.27%	9.40%	2.29	2.53
La, ppm	33.5	2.52	28.5	38.5	25.9	41.1	7.52%	15.04%	22.56%	31.8	35.2
Li, ppm	22.6	1.98	18.6	26.5	16.7	28.5	8.75%	17.51%	26.26%	21.5	23.7
Lu, ppb	73.1	13.3	46.5	99.7	33.2	113.0	18.19%	36.39%	54.58%	69.5	76.8
Mg, ppm	996	66	863	1128	797	1195	6.65%	13.30%	19.96%	946	1046
Mn, ppm	222	8	206	238	197	247	3.69%	7.39%	11.08%	211	233
Mo, ppm	5.22	0.471	4.28	6.17	3.81	6.64	9.02%	18.04%	27.06%	4.96	5.48
Na, wt.%	1.90	0.090	1.72	2.08	1.63	2.17	4.76%	9.52%	14.29%	1.81	2.00
Nb, ppm	14.4	1.04	12.3	16.5	11.3	17.5	7.22%	14.43%	21.65%	13.7	15.1
Nd, ppm	28.5	2.19	24.2	32.9	22.0	35.1	7.69%	15.37%	23.06%	27.1	30.0
Ni, ppm	6.54	0.67	5.19	7.88	4.52	8.56	10.30%	20.60%	30.90%	6.21	6.86
P, ppm	292	15	262	321	248	336	5.03%	10.07%	15.10%	277	306
Pb, ppm	318	16	287	349	271	365	4.89%	9.77%	14.66%	302	334
Pr, ppm	8.55	0.572	7.41	9.70	6.84	10.27	6.69%	13.38%	20.07%	8.13	8.98
Rb, ppm	98	3.5	91	105	87	108	3.62%	7.24%	10.86%	93	103
S, wt.%	1.50	0.036	1.43	1.57	1.39	1.60	2.40%	4.80%	7.20%	1.42	1.57
Sb, ppm	22.9	2.09	18.7	27.1	16.7	29.2	9.12%	18.23%	27.35%	21.8	24.1
Sc, ppm	3.77	0.239	3.29	4.25	3.05	4.48	6.34%	12.68%	19.02%	3.58	3.96
Se, ppm	10.6	1.1	8.4	12.8	7.3	13.9	10.40%	20.80%	31.20%	10.0	11.1
Sm, ppm	5.77	0.432	4.91	6.64	4.48	7.07	7.48%	14.97%	22.45%	5.48	6.06
Sn, ppm	3.36	0.177	3.00	3.71	2.83	3.89	5.26%	10.52%	15.78%	3.19	3.52
Sr, ppm	241	12	217	265	206	276	4.90%	9.80%	14.69%	229	253
Ta, ppm	1.11	0.076	0.96	1.26	0.88	1.34	6.85%	13.70%	20.55%	1.06	1.17
Tb, ppm	0.52	0.07	0.38	0.66	0.31	0.73	13.45%	26.90%	40.36%	0.49	0.55
Te, ppm	12.6	0.96	10.7	14.5	9.7	15.5	7.59%	15.19%	22.78%	12.0	13.2
Th, ppm	11.9	0.82	10.3	13.6	9.5	14.4	6.90%	13.80%	20.70%	11.3	12.5
Ti, wt.%	0.135	0.005	0.124	0.145	0.119	0.150	3.94%	7.89%	11.83%	0.128	0.141
Tl, ppm	1.44	0.092	1.26	1.63	1.17	1.72	6.39%	12.78%	19.18%	1.37	1.52
Tm, ppb	< 100	IND	IND	IND	IND	IND	IND	IND	IND	IND	IND
U, ppm	4.64	0.216	4.21	5.07	3.99	5.29	4.66%	9.32%	13.98%	4.41	4.87
V, ppm	12.1	0.88	10.4	13.9	9.5	14.7	7.24%	14.48%	21.73%	11.5	12.7
W, ppm	6.13	0.339	5.46	6.81	5.12	7.15	5.52%	11.04%	16.56%	5.83	6.44
Y, ppm	11.1	0.51	10.1	12.1	9.6	12.6	4.57%	9.14%	13.71%	10.5	11.7
Yb, ppm	0.54	0.06	0.43	0.65	0.38	0.71	10.14%	20.29%	30.43%	0.52	0.57
Zn, ppm	318	6	307	330	301	336	1.85%	3.70%	5.55%	302	334
Zr, ppm	186	11	165	207	154	218	5.67%	11.34%	17.01%	177	195

SI unit equivalents: ppm, parts per million  $\equiv$  mg/kg  $\equiv$   $\mu$ g/g  $\equiv$  0.0001 wt.%  $\equiv$  1000 ppb, parts per billion.

Note 1: intervals may appear asymmetric due to rounding.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.

Table 5 continued.

Constituent	Certified Value	Absolute Standard Deviations					Relative Standard Deviations			5% window	
		1SD	2SD Low	2SD High	3SD Low	3SD High	1RSD	2RSD	3RSD	Low	High
<b>Aqua Regia Digestion</b>											
Ag, ppm	50.0	2.86	44.3	55.7	41.4	58.6	5.72%	11.44%	17.16%	47.5	52.5
Al, wt. %	0.630	0.035	0.560	0.699	0.525	0.734	5.53%	11.05%	16.58%	0.598	0.661
As, ppm	276	17	242	310	225	327	6.16%	12.32%	18.48%	263	290
Be, ppm	0.47	0.041	0.39	0.55	0.35	0.60	8.60%	17.20%	25.80%	0.45	0.50
Bi, ppm	18.0	1.58	14.9	21.2	13.3	22.8	8.76%	17.52%	26.28%	17.1	18.9
Ca, wt. %	0.578	0.030	0.518	0.639	0.488	0.669	5.22%	10.44%	15.67%	0.549	0.607
Cd, ppm	2.04	0.177	1.68	2.39	1.50	2.57	8.71%	17.42%	26.13%	1.94	2.14
Ce, ppm	38.5	2.32	33.9	43.2	31.6	45.5	6.01%	12.03%	18.04%	36.6	40.5
Co, ppm	2.55	0.121	2.31	2.80	2.19	2.92	4.72%	9.44%	14.16%	2.43	2.68
Cr, ppm	24.6	2.26	20.1	29.1	17.8	31.4	9.17%	18.34%	27.50%	23.4	25.8
Cs, ppm	1.15	0.092	0.96	1.33	0.87	1.42	8.05%	16.09%	24.14%	1.09	1.20
Cu, wt. %	0.101	0.003	0.096	0.106	0.093	0.109	2.51%	5.03%	7.54%	0.096	0.106
Fe, wt. %	1.94	0.114	1.71	2.17	1.60	2.28	5.85%	11.70%	17.56%	1.84	2.04
Ga, ppm	3.77	0.189	3.39	4.15	3.20	4.34	5.02%	10.04%	15.06%	3.58	3.96
Ge, ppm	0.084	0.024	0.036	0.132	0.012	0.156	28.55%	57.10%	85.65%	0.080	0.088
Hf, ppm	1.11	0.107	0.89	1.32	0.78	1.43	9.71%	19.43%	29.14%	1.05	1.16
Hg, ppm	0.20	0.02	0.15	0.25	0.13	0.27	11.79%	23.58%	35.36%	0.19	0.21
In, ppm	0.42	0.016	0.39	0.45	0.37	0.47	3.89%	7.78%	11.68%	0.40	0.44
K, wt. %	0.246	0.019	0.208	0.285	0.188	0.304	7.83%	15.67%	23.50%	0.234	0.258
La, ppm	19.9	1.69	16.6	23.3	14.9	25.0	8.48%	16.96%	25.44%	18.9	20.9
Li, ppm	7.78	0.695	6.39	9.17	5.69	9.86	8.94%	17.88%	26.82%	7.39	8.16
Mg, ppm	411	20	371	451	351	472	4.90%	9.81%	14.71%	391	432
Mn, ppm	192	8	175	208	167	217	4.32%	8.64%	12.96%	182	201
Mo, ppm	4.83	0.281	4.27	5.40	3.99	5.68	5.81%	11.62%	17.42%	4.59	5.08
Na, wt. %	0.067	0.008	0.051	0.082	0.044	0.090	11.57%	23.14%	34.71%	0.064	0.070
Ni, ppm	6.39	0.419	5.56	7.23	5.14	7.65	6.56%	13.12%	19.67%	6.07	6.71
P, ppm	178	8	162	193	154	201	4.39%	8.78%	13.18%	169	187
Pb, ppm	234	14	205	262	191	276	6.08%	12.15%	18.23%	222	245
Rb, ppm	11.9	0.71	10.5	13.4	9.8	14.1	5.98%	11.96%	17.94%	11.3	12.5
S, wt. %	0.807	0.038	0.731	0.884	0.692	0.923	4.76%	9.51%	14.27%	0.767	0.848
Sb, ppm	18.0	2.2	13.6	22.4	11.4	24.6	12.22%	24.45%	36.67%	17.1	18.9
Sc, ppm	1.00	0.13	0.74	1.26	0.61	1.40	13.08%	26.17%	39.25%	0.95	1.05
Se, ppm	10.0	1.2	7.6	12.5	6.3	13.7	12.23%	24.45%	36.68%	9.5	10.5
Sn, ppm	1.19	0.112	0.97	1.42	0.86	1.53	9.33%	18.67%	28.00%	1.13	1.25
Sr, ppm	33.2	2.95	27.3	39.1	24.4	42.1	8.87%	17.74%	26.61%	31.6	34.9
Tb, ppm	0.28	0.06	0.16	0.40	0.10	0.45	21.20%	42.40%	63.59%	0.26	0.29
Te, ppm	13.0	0.83	11.3	14.6	10.5	15.5	6.41%	12.83%	19.24%	12.3	13.6
Th, ppm	6.96	0.73	5.50	8.42	4.77	9.15	10.49%	20.99%	31.48%	6.61	7.31
Ti, ppm	220	22	176	264	155	285	9.92%	19.84%	29.76%	209	231
Tl, ppm	1.08	0.084	0.91	1.25	0.82	1.33	7.82%	15.64%	23.46%	1.02	1.13
U, ppm	2.22	0.119	1.99	2.46	1.87	2.58	5.35%	10.70%	16.06%	2.11	2.34

SI unit equivalents: ppm, parts per million  $\equiv$  mg/kg  $\equiv$   $\mu$ g/g  $\equiv$  0.0001 wt. %  $\equiv$  1000 ppb, parts per billion.

Note 1: intervals may appear asymmetric due to rounding.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.

**Table 5 continued.**

Constituent	Certified Value	Absolute Standard Deviations					Relative Standard Deviations			5% window	
		1SD	2SD Low	2SD High	3SD Low	3SD High	1RSD	2RSD	3RSD	Low	High
<b>Aqua Regia Digestion continued</b>											
V, ppm	3.83	0.51	2.80	4.85	2.28	5.37	13.45%	26.90%	40.34%	3.63	4.02
W, ppm	1.86	0.184	1.49	2.22	1.31	2.41	9.89%	19.77%	29.66%	1.76	1.95
Y, ppm	5.34	0.528	4.28	6.39	3.76	6.92	9.88%	19.77%	29.65%	5.07	5.61
Yb, ppm	0.20	0.020	0.16	0.24	0.14	0.26	9.66%	19.31%	28.97%	0.19	0.21
Zn, ppm	267	7	252	282	244	289	2.79%	5.59%	8.38%	253	280
Zr, ppm	38.3	5.5	27.2	49.3	21.7	54.9	14.47%	28.93%	43.40%	36.3	40.2

SI unit equivalents: ppm, parts per million  $\equiv$  mg/kg  $\equiv$   $\mu$ g/g  $\equiv$  0.0001 wt.%  $\equiv$  1000 ppb, parts per billion.

Note 1: intervals may appear asymmetric due to rounding.

Note 2: the number of decimal places quoted does not imply accuracy of the certified value to this level but are given to minimise rounding errors when calculating 2SD and 3SD windows.

## PARTICIPATING LABORATORIES

1. AGAT Laboratories, Mississauga, Ontario, Canada
2. Alex Stewart International, Mendoza, Argentina
3. ALS, Johannesburg, South Africa
4. ALS, Lima, Peru
5. ALS, Loughrea, Galway, Ireland
6. ALS, Perth, WA, Australia
7. ALS, Vancouver, BC, Canada
8. American Assay Laboratories, Sparks, Nevada, USA
9. ANSTO, Lucas Heights, NSW, Australia
10. ARGETEST Mineral Processing, Ankara, Central Anatolia, Turkey
11. Bureau Veritas Commodities Canada Ltd, Vancouver, BC, Canada
12. Bureau Veritas Geoanalytical, Perth, WA, Australia
13. Bureau Veritas Minerals, Hermosillo, Sonora, Mexico
14. Inspectorate (BV), Lima, Peru
15. Intertek Genalysis, Perth, WA, Australia
16. Intertek Testing Services, Townsville, QLD, Australia
17. Intertek Testing Services Philippines, Cupang, Muntinlupa, Philippines
18. PT Geoservices Ltd, Cikarang, Jakarta Raya, Indonesia
19. PT Intertek Utama Services, Jakarta Timur, DKI Jakarta, Indonesia
20. PT SGS Indo Assay Laboratories, Jakarta, Indonesia
21. SGS, Ankara, Anatolia, Turkey
22. SGS de Mexico SA de CV, Cd. Industrial, Durango, Mexico
23. SGS del Peru, Lima, Peru
24. SGS Lakefield Research Ltd, Lakefield, Ontario, Canada
25. Shiva Analyticals Ltd, Bangalore North, Karnataka, India


**Please note: Above numbered alphabetical list of participating laboratories does not reflect the Lab ID numbering on the scatter plots below.**



# Figure 1. Au by Fire Assay in OREAS 601b

SPC.1399.OREAS600b Mt Carlton.OREAS 601b.4.Fire Assay.Au.Lab.190219.143553.SN

ppm



Legend:

- Inlier (Blue circle)
- INAA (Black diamond)
- Indiv. Outlier (Red circle)
- Batch Outlier (Red circle)
- Rec. Value (Green line)
- ± 5% (Yellow line)
- ± 3SD (Magenta line)

Lab

Figure 2. Ag by 4-Acid in OREAS 601b

SPC.1399.OREAS600b Mt Carlton.OREAS 601b.4.4-Acid.Ag.Lab.190219.143357.SN





Figure 3. Cu by 4-Acid in OREAS 601b

SPC.1399.OREAS600b Mt Carlton.OREAS 601b.4.4-Acid.Cu.Lab.190219.143434.SS



## PREPARER AND SUPPLIER

Certified reference material OREAS 601b was prepared, certified and supplied by:



ORE Research & Exploration Pty Ltd  
37A Hosie Street  
Bayswater North VIC 3153  
AUSTRALIA

Tel: +613-9729 0333  
Fax: +613-9729 8338  
Web: [www.ore.com.au](http://www.ore.com.au)  
Email: [info@ore.com.au](mailto:info@ore.com.au)

It is packaged under nitrogen in 10g and 60g units in laminated foil pouches and 500g units in plastic jars.

## METROLOGICAL TRACEABILITY

The analytical samples were selected in a manner to represent the entire batch of prepared CRM. This 'representivity' was maintained in each submitted laboratory sample batch and ensures the user that the data is traceable from sample selection through to the analytical results that underlie the consensus values. Each analytical data set has been validated by its assayer through the inclusion of internal reference materials and QC checks during analysis.

The laboratories were chosen on the basis of their competence (from past performance in inter-laboratory programs undertaken by ORE Pty Ltd) for a particular analytical method, analyte or analyte suite, and sample matrix. Most of these laboratories have and maintain ISO 17025 accreditation. The certified values presented in this report are calculated from the means of accepted data following robust statistical treatment as detailed in this report.

Guide ISO/TR 16476:2016, section 5.3.1 describes metrological traceability in reference materials as it pertains to the transformation of the measurand. In this section it states, *"Although the determination of the property value itself can be made traceable to appropriate units through, for example, calibration of the measurement equipment used, steps like the transformation of the sample from one physical (chemical) state to another cannot. Such transformations may only be compared with a reference (when available), or among themselves. For some transformations, reference methods have been defined and may be used in certification projects to evaluate the uncertainty associated with such a transformation. In other cases, **only a comparison among different laboratories using the same method is possible. In this case, certification takes place on the basis of agreement among independent measurement results** (see ISO Guide 35:2006, Clause 10)."*

## COMMUTABILITY

The measurements of the results that underlie the certified values contained in this report were undertaken by methods involving pre-treatment (digestion/fusion) of the sample. This served to reduce the sample to a simple and well understood form permitting calibration using simple solutions of the CRM. Due to these methods being well understood and highly effective, commutability is not an issue for this CRM. All OREAS CRMs are sourced from natural ore minerals meaning they will display similar behaviour as routine 'field' samples in the relevant measurement process. Care should be taken to ensure 'matrix matching' as close as practically achievable. The matrix and mineralisation style of the CRM is described in the 'Source Material' section and users should select appropriate CRMs matching these attributes to their field samples.

## INTENDED USE

OREAS 601b is intended to cover all activities needed to produce a measurement result. This includes extraction, possible separation steps and the actual measurement process (the signal producing step). OREAS 601b may be used to calibrate the entire procedure by producing a pure substance CRM transformed into a calibration solution. OREAS 601b is intended for the following uses:

- For the monitoring of laboratory performance in the analysis of analytes reported in Table 1 in geological samples;
- For the verification of analytical methods for analytes reported in Table 1;
- For the calibration of instruments used in the determination of the concentration of analytes reported in Table 1.

## STABILITY AND STORAGE INSTRUCTIONS

OREAS 601b has been prepared from sulphide bearing ores blended with rhyodacite. It contains a small proportion of reactive sulphides (~1.5% S) and as a precaution to prevent oxidation, has been packaged under a nitrogen environment in single-use laminated foil pouches (10g and 60g units). In its unopened state and under normal conditions of storage the CRM has a shelf life beyond ten years. Its stability will be monitored at regular intervals and purchasers notified if any changes are observed.

## INSTRUCTIONS FOR CORRECT USE

The certified values for OREAS 601b refer to the concentration level in its packaged state. It should not be dried prior to weighing and analysis.

## HANDLING INSTRUCTIONS

Fine powders pose a risk to eyes and lungs and therefore standard precautions such as the use of safety glasses and dust masks are advised.

## LEGAL NOTICE

Ore Research & Exploration Pty Ltd has prepared and statistically evaluated the property values of this reference material to the best of its ability. The Purchaser by receipt hereof releases and indemnifies Ore Research & Exploration Pty Ltd from and against all liability and costs arising from the use of this material and information.

## DOCUMENT HISTORY

Revision No.	Date	Changes applied
2	20 <sup>th</sup> June 2019	Corrected 'ppm' to 'wt.%' for Cu in Summary Statistics (page 1).
1	27 <sup>th</sup> February 2019	Corrected 'Summary Statistics for Key Analytes' table.
0	21 <sup>st</sup> February 2019	First publication.

## QMS ACCREDITED

ORE Pty Ltd is accredited to ISO 9001:2015 by Lloyd's Register Quality Assurance Ltd for its quality management system including development, manufacturing, certification and supply of CRMs.



## CERTIFYING OFFICER

A handwritten signature in black ink, appearing to read 'S.H.', is positioned above the name of the certifying officer.

12<sup>th</sup> June, 2019

---

Craig Hamlyn (B.Sc. Hons - Geology), Technical Manager - ORE P/L

## REFERENCES

Govett, G.J.S. (1983), ed. Handbook of Exploration Geochemistry, Volume 2: Statistics and Data Analysis in Geochemical Prospecting (Variations of accuracy and precision), P.O. Box 330, 1000 AH Amsterdam, The Netherlands.

Ingamells, C. O. and Switzer, P. (1973), Talanta 20, 547-568.

ISO Guide 30:2015, Terms and definitions used in connection with reference materials.

ISO Guide 31:2015, Reference materials – Contents of certificates and labels.

ISO Guide 35:2017, Certification of reference materials - General and statistical principals.

ISO 16269:2014, Statistical interpretation of data – Part 6: Determination of statistical tolerance intervals.

Munsell Rock Color Book (2014), Rock-Color Chart Committee, Geological Society of America (GSA), 4300 44th Street SE, Grand Rapids, MI 49512.